



Tinytag View 2s are all housed in an attractive IP65 case and have an integral display. All feature high

reading accuracy and resolution, large memories, a fast offload speed and a low battery monitor.

With an integral stainless steel probe the TV-4076

is ideal for measuring product core temperatures.

Tinytag View 2 Temperature Logger with Integral Stab Probe (-30 to +50 ℃)

TV-4076

Issue 7 9th August 2019 E&OE

VEW 2

Popular Applications

- Product Production and Storage
- Product Core Temperatures

Features

- Temperature recorder
- LCD display of current readings
- 30,000 reading capacity
- High Accuracy
- High Reading Resolution
- Fast Data Offload
- 2 user-programmable alarms
- Low battery monitor
- User-replaceable battery





Tinytag View 2 Temperature Logger with Integral Stab Probe (-30 to +50 °C) TV-4076

Issue 7: 9th August 2019 (E&OE)



ea	tures	


ading Capacity type	30,000 readings Non Volatile 4 digits + indicators
lodes	°C or °F
Refresh Rate	Every 2 seconds
start	Magnetic Switch
Start	Relative / Absolute
	(up to 45 days)
ions	When full
	After n Readings
	Never (overwrite oldest data)
Types	Actual, Min, Max
Interval	1 sec to 10 days
	While stopped or when
	logging in minutes
	mode
	2 fully programmable; latchable

Reading Specification

Reading Range Sensor Type **Response Time** Logger Resolution **Display Resolution Temperature Stability**

-30 ℃ to +50 ℃ (-22 ℃ to 122 ℃) 10K NTC Thermistor 15 seconds to 90% FSD in water 0.02 ℃ or better 0.1 ℃ or 0.1 °F 0.005℃/℃ Change from 25℃

Logger Accuracy



Physical Specification

IP Rating Operational Range Case Dimensions Diameter Length (inc. prob Width Depth Weight

*	IP65 splash proof (see notes) -30℃ to +70℃
	60mm / 2.36"
oe)	150mm / 5.91"
	77mm / 3.03"
	35mm / 1.38"
	95g / 3.35oz

*The Operational Range indicates the physical limits to which the unit can be exposed, not the reading range over which it will record

Notes

The battery fitted in this product is a single cell containing less that 1g of lithium and meets the requirements of the UN Manual of Tests and Criteria, Part III, Subsection 38.3.

Recommended Battery Types

SAFT LS14250. Tekcell SBAA02P or Eve ER14250

The logger will operate with other 1/2AA 3.6V Lithium batteries but performance cannot be guaranteed.

Replacement Interval Annually

Before replacing the battery the data logger must be stopped.

After removing an old battery from a logger, wait five minutes before inserting the new one.

Data stored on the logger will be retained after a battery is replaced.

The clarity of the display may change at extremes of temperature.

If used at low temperatures the data logger should be allowed to warm to room temperature before it is opened to avoid condensation forming inside the unit.

The IP65 rating is valid only when the unit's connector cap is securely fitted.

Calibration

This unit is configured to meet Gemini's quoted accuracy specification during its manufacture

We recommend that the calibration of this unit should be checked annually against a calibrated reference meter.

A certificate of calibration, traceable to a national standard, can be supplied for an additional charge either at the point of purchase, or if the unit is returned for a service calibration.

Approvals

Gemini Data Loggers (UK) Ltd. operates a Business Management System which conforms to ISO 9001 and ISO 14001.



Required and Related Products

To use this data logger you will require the following software:

SWCD-0040: Tinytag Explorer software

and a

CAB-0007-USB: Tinytag Ultra/Plus/View USB Download Cable

The SWCD-0040 software and CAB-0007-USB cable can be ordered together in a pack using the part number SWPK-7-USB.

Further Related Products

SER-9500: Tinytag Data Logger Service Kit ACS-6000: Trigger Start Magnet

www.tinytag.info